

Journal section: *New technologies in Dentistry*
Publication Types: *Research*

doi:10.4317/jced.51230
<http://dx.doi.org/10.4317/jced.51230>

A comparative assessment of the accuracy of electronic apex locator (Root ZX) in the presence of commonly used irrigating solutions

Osama Khattak¹, Ebadullah Raidullah², Maria-Louis Francis³

¹ Assistant Professor of Endodontics and Co Ordinator Dental Clinics, RAK College of Dental Sciences, Ras al Khaimah, UAE

² Junior Instructor, RAK College of Dental Sciences, Ras al Khaimah, UAE

³ Intern dentist, RAK College of Dental Sciences, Ras al Khaimah, UAE

Correspondence:
RAK College of Dental Sciences
P.O Box Number 12973
Ras Al Khaimah
United Arab Emirates
khattakosama@gmail.com

Received: 28/07/2013
Accepted: 29/09/2013

Khattak O, Raidullah E, Francis ML. A comparative assessment of the accuracy of electronic apex locator (Root ZX) in the presence of commonly used irrigating solutions. J Clin Exp Dent. 2014;6(1):e41-6. <http://www.medicinaoral.com/odo/volumenes/v6i1/jcedv6i1p41.pdf>

Article Number: 51230 <http://www.medicinaoral.com/odo/indice.htm>
© Medicina Oral S. L. C.I.F. B 96689336 - eISSN: 1989-5488
eMail: jced@jced.es
Indexed in:
Scopus
DOI® System

Abstract

Objectives: This study aimed to evaluate the accuracy of Root ZX in determining working length in presence of normal saline, 0.2% chlorhexidine and 2.5% of sodium hypochlorite.

Material and Methods: Sixty extracted, single rooted, single canal human teeth were used. Teeth were decoronated at CEJ and actual canal length determined. Then working length measurements were obtained with Root ZX in presence of normal saline 0.9%, 0.2% chlorhexidine and 2.5% NaOCl. The working length obtained with Root ZX were compared with actual canal length and subjected to statistical analysis.

Results: No statistical significant difference was found between actual canal length and Root ZX measurements in presence of normal saline and 0.2% chlorhexidine. Highly statistical difference was found between actual canal length and Root ZX measurements in presence of 2.5% of NaOCl, however all the measurements were within the clinically acceptable range of ± 0.5 mm.

Conclusion: The accuracy of EL measurement of Root ZX within ± 0.5 mm of AL was consistently high in the presence of 0.2% chlorhexidine, normal saline and 2.5% sodium hypochlorite.

Clinical significance: This study signifies the efficacy of ROOT ZX (Third generation apex locator) as a dependable aid in endodontic working length.

Key words: *Electronic apex locator, working length, root ZX accuracy, intracanal irrigating solutions.*

Introduction

The removal of all pulp tissue, necrotic material and microorganisms from the root canal is essential for endodontic success. On the other hand determination of an accurate working length is also a critical step in endodontic therapy (1). Therefore proper instrumentation upto the apical constriction or also called as the cemento-dentinal junction (2) as seen earlier is also one of the vital factor for a good prognosis. Traditionally, the point of termination for endodontic instrumentation and obturation has been determined by taking radiographs. But just how accurate is this radiographic measurement? For one thing, accuracy depends on the radiographic technique used. Forsberg, in Norway, demonstrated that paralleling technique was “significantly more reliable” than the bisecting-angle technique (3). A US Army group, however, found that the paralleling technique was absolutely accurate only 82% of the time (4) One study suggested that paralleling technique magnified actual tooth length by 5.4% (5). As Olson et al. pointed out, 82 to 89% accuracy is not 100%, so they recommended back-up methods such as tactile feel, moisture on the tip of a paper point, or electronic apex locators (4). Similar results and recommendations have been reported worldwide (6-11).

An *in vivo* histological study found that the most favourable histological conditions were when the instrumentation and obturation remained short of the apical constriction and that extruded gutta-percha and sealer always caused a severe inflammatory reaction despite the absence of pain (12). The problem clinicians face is how to accurately identify and prepare to this landmark – the ‘working length’ – and achieve maximum success.

The radiographs are definitely supportive for the instrumentation upto the apical constriction, but they can also prove deceptive due to improper angulation of the cone. The radiographic assessment technique is sensitive in both its exposure and interpretation (13). Also, the image obtained is a two-dimensional image of a three dimensional object. Therefore keeping the above limitations in mind an alternative for radiographic apex locator was very much needed. First electronic apex locator was introduced in 1918 by Custer et al. His ideas were later revisited by Suzuki in 1942 (14) and Sunada in 1962 (15) for the invention of the modern electronic apex locator apex. Thus the development of modern electronic apex locator has helped make the assessment of working length more accurate and predictable (16).

All apex locators function by using the human body to complete an electrical circuit. One side of the apex locator’s circuitry is connected to an endodontic instrument. The other side is connected to the patient’s body, either by a contact to the patient’s lip or by an electrode held in the patient’s hand. The electrical circuit is com-

plete when the endodontic instrument is advanced apically inside the root canal until it touches periodontal tissue. The display on the apex locator indicates that the apical area has been reached. Many *in vitro* accuracy studies were conducted on models using an extracted tooth in an electrolyte to simulate clinical conditions. The ideal conditions in *in vitro* testing may give accuracy results higher than those obtainable in clinical practice (17-25).

Therefore the main aim of our study was to test the efficacy of ROOT ZX (J. Morita Mfg Corp., Kyoto, Japan), a third-generation apex locator based on “ratio method” that uses dual frequency and comparative impedance principles (26,27). The Root ZX is mainly based on detecting the change in electrical capacitance that occurs near the apical constriction. Some of the advantages of the Root ZX are that it requires no adjustment or calibration and can be used when the canal is filled with strong electrolyte or when the canal is “empty” and moist (26).

Material and Methods

Sixty, straight, single-rooted permanent human teeth with mature apices were selected for this study. The teeth were cleaned of calculus, soft tissues, and debris with hand instrumentation and stored in distilled water until used. The type I canal configuration was confirmed by using digital radiograph (Gendex, Dentsply) in mesiodistal and labiolingual planes. Teeth with resorption, curvatures, open apices, or radiographically invisible canals were excluded from the study.

The teeth were decoronated at the level of cementoenamel junction with a diamond disc to allow access to the root canal and to provide a stable reference for all measurements. The coronal portion of each canal was preflared using sequential Gates Glidden drills #4, #3, and #2 (Mani Inc., Japan), irrigated with saline and pulp extirpated with a barbed broach (Spirocolorinox, Dentsply).

- Measurement of actual working length

Teeth were numbered 1–60 and the actual canal length (AL) was determined by introducing a size 10 or 15 k-file (Mani Inc., Japan) into the canal until its tip emerged through the major apical foramen. 3.5x magnification loops were used during this procedure for enhanced visibility. The long axis of the tooth was placed perpendicular to the line of sight and the tip of the file will be positioned tangential to the major apical foramen (28,29). After carefully adjusting the silicone stopper to the reference point, the file was withdrawn from the root canal, and the distance between the file tip and silicone stopper was measured with a digital caliper (Mitutoyo Co., Japan) and the reading was noted down.

To simulate the periodontium, this study used the *in vitro* model as designed by Donnelly (30). A polystyrene

specimen bottle (40 ml) was filled with warmed gelatin solution and refrigerated for 2 h to allow gelatin to set. The apical two-third of the root was embedded in gelatin, and the tooth was stabilized to the lid of a container with auto-polymerizing resin as described by Higa *et al* (29). The lip electrode was also placed in gelatin through another opening in the lid.

- Working model for electronic working length determination

The irrigants tested will be: 0.9% saline, 2.5% sodium hypochlorite (NaOCl) and 0.2% chlorhexidine gluconate (CHX) (*Curasept, Italy*). The irrigant to be tested was introduced into the canal with a 23-gauge needle.

The EAL tested in this experiment was: Root ZX (J. Morita Mfg Corp., Kyoto, Japan). It was used according to manufacturer’s instructions (31). Depending on the size of the canal, #15 or #20 K-file (Mani Inc., Japan), was attached to the file holder and introduced into the canal. The meter’s 0.5 mm reading was set between the “APEX” and “1” (factory setting) as indicated by a flashing bar and will be used for electronic measurements. The file was gently inserted into the root canal until the “APEX” signal will be displayed. The file was then gently retracted until the display showed a flashing image of the root canal and a flashing bar between APEX and 1 (0.5 reading). The silicone stopper on the file will be carefully adjusted to a reference point, and the file was withdrawn to measure the distance between the silicone stopper and the file tip and noted down. This was recorded as the electronically measured canal length (EL).

The electronically measured Working length (EL) was compared with the actual canal (AL) length measured conventionally and scores were attributed to the resulting values (32) (Table 1).

The canal length was assessed for each tooth with individual irrigants. To prevent cross-contamination: (a) fresh gelatin was used for the individual irrigant, and (b) the root canals was irrigated with ethanol and dried with paper points. The results obtained (in millimeters) were recorded. The difference between the median of electronically measured length (EL) and the AL was calculated for each tooth in the presence of all irrigating solutions. The resulting difference in working length was noted down.

A paired t-test was employed to statistically analyze the significance of mean difference between EL and AL. One-way ANOVA was also employed along with Welch & Brown-Forsythe analysis to assess the significance of difference among various irrigants in their estimations of the canal length. Level of significance was set at P <0.05. The analysis was performed with Statistical and Presentational System Software (SPSS 20.0, SPSS Inc, Chicago, IL).

Results

Table 2, Table 3 and Table 4 shows the actual working length obtained with conventional method and the electronically measured working length obtained with Root Zx with different group of irrigants (*0.9% Saline, 2.5% NaOCl & 0.2% Chlorohexidine*). It also indicates the score (*as mentioned in table 1*) and the difference in working length obtained per tooth.

Table 2. Group 1 (0.9% Saline as irrigant).

Tooth Number	Actual length (AL) with Conventional Method (mm)	Electronically Calculate (EL) Working length with Root ZX (mm)	Score	Difference (AL-EL)
1	23	23	0	0
2	23	23	0	0
3	20	20	0	0
4	22	22	0	0
5	21.5	21.5	0	0
6	23.5	23.5	0	0
7	21	21	0	0
8	21	21	1	0
9	22	21.6	1	0.4
10	23	22.6	1	0.4
11	24	24	0	0
12	20	19.6	1	0.4
13	20	20	0	0
14	22	22	0	0
15	20	20	0	0
16	19	19	0	0
17	20	20	0	0
18	20	19.6	1	0.4
19	23	23	0	0
20	22	22	0	0

Table 1. Comparison between the electronically measured Working length (EL) and the actual canal (AL).

Score	Situation
0	Working Length (WL) equal to gold standard
1	Working Length (WL) from 0.5 to 1mm shorter than gold standard
2	Working Length (WL) > 1mm shorter than gold standard
3	Working Length (WL) exceeds gold standard

Table 3. Group 1 (2.5% Sodium Hypochlorite - NaOCl).

Tooth Number	Actual length (AL) with Conventional Method (mm)	Electronically Calculate (EL) Working length with Root ZX (mm)	Score	Difference (AL-EL)
21	23	23	0	0
22	20.5	20	1	0.5
23	22	22	0	0
24	21	20.6	1	0.4
25	22.5	22.5	0	0
26	21.5	21.1	1	0.4
27	21	20.7	1	0.3
28	23	23	0	0
29	23	23	0	0
30	23	22.5	1	0.5
31	22	22	0	0
32	21	21	0	0
33	22	22	0	0
34	23	22.6	1	0.4
35	21	21	0	0
36	23	22.5	1	0.5
37	23	23	0	0
38	21	21	0	0
39	23	22.6	1	0.4
40	21	21	0	0

Table 4. Group 1 (0.2% Chlorohexidine gluconate - CHX)

Tooth Number	Actual length (AL) with Conventional Method (mm)	Electronically Calculate (EL) Working length with Root ZX (mm)	Score	Difference (AL-EL)
41	19	18.8	1	0.2
42	19	19	0	0
43	18	18	0	0
44	18	18	0	0
45	22	21.7	1	0.3
46	21	21	0	0
47	23	23	0	0
48	19	19	0	0
49	19	19	0	0
50	20	20	0	0
51	23	23	0	0
52	18	18	0	0
53	21	21	0	0
54	19	19	0	0
55	20	20	0	0
56	20	20	0	0
57	21	21	0	0
58	23	23	0	0
59	22	22	0	0
60	21	21	0	0

The mean values of actual canal length and electronically measured working length with Root ZX along with their differences for all three groups are given in Table 5.

Table 5. Mean (SD) of AL and EL measurements with Root ZX in the presence of various irrigants.

Irrigants	Actual Length Mean (SD)	Electronic Length Root ZX Mean (SD)	Difference
2.5% NaOCl	22.025 (0.9386)	21.855 (0.9600)	0.17 (0.218)
0.2% CHX	20.300 (1.6890)	20.275 (1.6830)	0.03 (0.079)
0.9% Saline	21.500 (1.4600)	21.420 (1.4835)	0.08 (0.164)

n=20 for each irrigant; AL = Actual length; EL = Electronically measured canal Length; SD = Standard Deviation; NaOCl = Sodium Hypochlorite; CHX = Chlorohexidine

Table 6 shows that the working length differences obtained by different groups of irrigants is statistically significant as P value = 0.024. Robust test of equality of means was also employed and the outcome of Welch and Brown-Forsythe analysis is shown in Table 7 where P value is 0.022 and 0.026 respectively. Therefore proving further statistical significance.

Our study indicated that the differences in working length obtained with 2.5% NaOCl were significantly larger than those obtained with 0.2% Chlorohexidine and 0.9% Saline. Although statistically significant differences existed between the irrigants, the majority of the readings were within the acceptable range of ±0.5 mm for Root ZX

The results indicate that root ZX was accurate within ±0.5mm 100% of the time with all the three test irrigants.

Table 6. Anova.

Difference					
	Sum of Squares	df	Mean Square	F	Sig.
Between Groups	.214	2	.107	3.989	.024
Within Groups	1.532	57	.027		
Total	1.746	59			

Table 7. Robust Test of equality of means.

Difference				
	Statistic ^a	df1	df2	Sig.
Welch	4.282	2	32.098	.022
Brown-Forsythe	3.989	2	40.901	.026

a. Asymptotically F distributed.

Discussion

The first-generation EALs were resistance-based and the second-generation EALs were impedance-based apex locators (33). The main shortcomings of these EALs included poor accuracy in the presence of fluids and pulp tissue, and the need for calibration (34). The frequency-based third-generation EALs have more powerful microprocessors and are able to process mathematical quotient and algorithm calculations required

to give accurate readings (35). Root ZX (J. Morita Mfg Corp., Kyoto, Japan) is a third-generation EAL that uses dual frequency and comparative impedance principle is based on the “ratio method” for measuring canal length. This method simultaneously measures the impedance values at two frequencies (8 and 0.4 kHz) and calculates a quotient of impedances. This quotient is expressed as a position of the file in the canal. Root ZX requires no calibration, and can be used when the canal is filled with a strong electrolyte (36).

The present study used an *in vitro* model as described by Donnelly (30) to obtain accurate measurements. The advantages of the model were its simplicity, ease of use and the ability to have strict control over the tested experimental condition. A disadvantage of the model is its inability to fully simulate *in vivo* conditions. Huang supported the use of *in vitro* models for evaluation of electronic apex locators (37). According to him, when tip of the file passes through narrow apical foramen, the physical properties of the foramen itself produce the electrical resistance gradient.

The use of irrigating solutions is an important aspect of endodontic treatment. Several studies using advanced techniques such as micro-computed tomography (CT) scanning have demonstrated that proportionally large areas of the main root-canal wall remain untouched by the instruments (38). Sodium hypochlorite (NaOCl) is the most popular irrigating solution. NaOCl ionizes in water into Na and the hypochlorite ion, OCl, establishing an equilibrium with hypochlorous acid (HOCl). At acidic and neutral pH, chlorine exists predominantly as HOCl, whereas at high pH of 9 and above, OCl predominates (39). Hypochlorous acid is responsible for the antibacterial activity. NaOCl is commonly used in concentrations between 0.5% and 6%. It is a potent antimicrobial agent, killing most bacteria instantly on direct contact. It also effectively dissolves pulpal remnants and collagen, the main organic components of dentin. Hypochlorite is the only root-canal irrigant of those in general use that dissolves necrotic and vital organic tissue. It is difficult to imagine successful irrigation of the root canal without hypochlorite.

Chlorhexidine gluconate (CHX) is widely used in disinfection in dentistry because of its good antimicrobial activity (40). CHX is marketed as a water-based solution and as a gel (with Natrosol). Some studies have indicated that the CHX gel has a slightly better performance than the CHX liquid but the reasons for possible differences are not known (41). CHX solutions in concentrations of 0.2–2% are considered toxicologically safe (42). However, there is paucity of research regarding the accuracy of EAL in presence of Chlorhexidine (43). It is imperative that the clinician should be confident of the fact that irrigating solution is not effecting the accurateness of the apex locator.

The results of this study showed that, Root ZX is 100% accurate within 0.5 mm from the apical foramen. The ± 0.5 mm to the foramen range has been considered as the strictest acceptable range (44). Thus, measurements attained within this tolerance are considered highly accurate. The results of measurements which are within this range are considered accurate. According to this study, statistically significant difference was found when measurements were done in canals irrigated with 2.5% sodium hypochlorite. However these were within the clinically acceptable range of within 0.5mm. The possible reason for this variation could be the higher electrical conductivity of sodium hypochlorite.

The use of electronic devices to determine working length has gained increasing popularity in recent years (45). The use of apex locators is very useful in scenarios like anatomical limitations for example zygomatic arch or maxillary sinus where at times. Also at times distortion of radiographs limits the accuracy of working length determination.

The results of the studies confirm that EALs can accurately determine the canal length within ± 0.5 mm from the apical constriction (46). The Root ZX has become a benchmark to which other devices are compared with. The results of this study are comparable to previously studies done (47-48).

This study shows that Root ZX can reliably be used for determining the position of the apical foramen in the presence of above mentioned irrigating solutions.

Conclusion

The results of the present study confirm that EALs can accurately determine the root canal length within ± 0.5 mm from the apical constriction. Therefore it is clinically safe and accurate to use Electronic apex locator with all the three mentioned irrigating solutions in this study.

References

1. Ingle JI, Himel VT, Hawrish CE, Glickman GE. Endodontic cavity preparation. In: Ingle IJ, Bakland LK, editors. Endodontics. 5th ed. Elsevier India: B.C. Decker; 2003. pp. 405-70.
2. Grove CJ The value of the dentinocemental junction in pulp canal surgery. *Journal of Dental Research*. 1931;11:466-8.
3. Forsberg J. Radiographic reproduction of endodontic “working length” comparing the paralleling and bisecting-angle techniques. *Oral Surg*. 1987;64:353.
4. Olson AK, Goerig AC, et al. The ability of the radiograph to determine the location of the apical foramen. *Int Endod J*. 1991;24:28.
5. Vande Voorde H, Bjorndahl A. Estimating endodontic “working length” with paralleling radiographs. *Oral Surg*. 1969;27:106.
6. Cox VS, Brown CE, Bricker SL. Radiographic interpretation of endodontic file length. *Oral Surg*. 1972;72:340.
7. Eckerboom M, Magnusson T. Evaluation of technical quality of endodontic treatment—reliability of intra-oral radiographs. *Endod Dent Traumatol*. 1997;13:259.
8. Bhakdinarenk A, Manson-Hing LR. Effect of radiographic technique upon the prediction of tooth length in intraoral radiography. *Oral Surg*. 1981;51:100.
9. Teo CS, Chan NC, Loh HS. The position of the apical foramen of the permanent incisors. *Aust Dent J*. 1988;33:51.

10. Mandel E. Obturation canalaire et position du foramen apical. *Rev Franc Endod.* 1983;31:49.
11. Shearer AC, Horner K, Wilson NHF. Radiovisography for length estimation in root canal treatment: an in vitro comparison with conventional radiography. *Int Endod J.* 1991;24:233.
12. Rieucci D, Langeland K. Apical limit of root canal instrumentation and obturation, part 2. A histological study. *Int Endod J.* 1998;31:394-409.
13. Gordon MPJ, Chandler NP. Electronic apex locators. *Int Endod J.* 2004;37:425-37.
14. Suzuki K. Experimental study on iontophoresis. *Japanese Journal of Stomatology.* 1942;16:411-29.
15. Sunada I. New method for measuring the length of the root canal. *Journal of Dental Research.* 1962;41:375-87.
16. Fouad AF, Reid LC. Effect of using electronic apex locators on selected endodontic treatment parameters. *J Endod.* 2000;26:364-7.
17. Aurelio JA, Nahmias Y, Gerstein H. A model for demonstrating an electronic canal length measuring device. *JOE.* 1983;9:568.
18. Nahmias Y, Aurelio JA, Gerstein H. An in vitro model for evaluation of electronic root canal measuring devices. *JOE.* 1987;13:209.
19. Donnelly JC. A simplified model to demonstrate the operation of an electronic root canal measuring devices. *JOE.* 1993;19:579.
20. Czerw RJ, Fulkerson MS, Donnelly JC. An in vitro test of a simplified model to demonstrate the operation of an electronic root canal measuring device. *JOE.* 1994;20:605.
21. Fouad AF, Krell KV. An in vitro comparison of five root canal length measuring instruments. *JOE.* 1989;15:577.
22. Kaufman AY, Katz A. Reliability of Root ZX apex locator tested by an in vitro model [abstract]. *JOE.* 1993;19:201.
23. Falchetta M, Castellucci A. In vitro evaluation and clinical impressions of the electronic apex locator Root ZX. *G Ital Endod.* 1993;4:173.
24. Kaufman AY, Fuchs M, Freidman S, et al. An in vitro model for practicing the use of nickel titanium files with combined automated engine devices using a built-in apex locator. [In preparation].
25. Olson AK, Goerig AC, Cavataio RE, Luciano J. The ability of the radiograph to determine the location of the apical foramen. *Int Endod J.* 1991;24:28.
26. Kobayashi C, Okiji T, Kawashima N, et al. A basic study on the electronic root canal length measurement. Part 3. Newly designed electronic root canal length measuring device using division method. *Jpn J Conserv Dent.* 1991;34:1442.
27. Kobayashi C, Suda H. New electronic canal measuring device based on the ratio method. *JOE.* 1994;20:111.
28. Chopra V, Grover S, Prasad SD. In vitro evaluation of the accuracy of two electronic apex locators. *J Conserv Dent.* 2008;11:82-5.
29. Higa RA, Adorno CG, Ebrahim AK, Suda H. Distance from file tip to the major apical foramen in relation to the numeric meter reading on the display of three different electronic apex locators. *Int Endod J.* 2009;42:1065-70.
30. Donnelly JC. A simplified model to demonstrate the operation of electronic root canal measuring devices. *J Endod.* 1993;19:579-80.
31. J Morita corp. Fully automatic root canal length measuring device. Root ZX operation instructions. 2000:1-9.
32. Menezes JVN. Estudo in vitro do grau de fidelidade de técnicas de odontometria em molares deciduos [Master's thesis]. Florianópolis: Universidade Federal de Santa Catarina; 1999.
33. Ingle JJ, Himel VT, Hawrish CE, Glickman GE. Endodontic cavity preparation. In: Ingle JJ, Bakland LK, editors. *Endodontics.* 5th ed. Elsevier India: B.C. Decker; 2003. pp. 405-570.
34. Kim E, Lee SJ. Electronic apex locator. *Dent Clin North Am.* 2004;48:35-54.
35. Gordon MPJ, Chandler NP. Electronic apex locators. *Int Endod J.* 2004;37:425-37.
36. Ingle JJ, Himel VT, Hawrish CE, Glickman GE. Endodontic cavity preparation. In: Ingle JJ, Bakland LK, editors. *Endodontics.* 5th ed. Elsevier India: B.C. Decker; 2003. pp. 405-570.
37. Huang L. An experimental study of the principle of electronic root canal measurement. *J Endodon.* 1987;13:60-4.
38. Peters OA, Schonenberger K, Laib A. Effects of four Ni-Ti preparation techniques on root canal geometry assessed by micro computed tomography. *Int Endod J.* 2001;34:221-30.
39. Barrette WC Jr, Hannum DM, Wheeler WD, et al. General mechanism for the bacterial toxicity of hypochlorous acid: abolition of ATP production. *Biochemistry.* 1989;28:9172-8.
40. Russell AD, Day MJ. Antibacterial activity of chlorhexidine. *J Hosp Infect.* 1993;25:229-38.
41. Ferraz CC, Gomes BP, Zaia AA, et al. In vitro assessment of the antimicrobial action and the mechanical ability of chlorhexidine gel as an endodontic irrigant. *J Endod.* 2001;27:452-5.
42. Yesilsoy C, Whitaker E, Cleveland D, Phillips E, Trope M. Antimicrobial and toxic effects of established and potential root canal irrigants. *J Endod.* 1995;21:513-5.
43. Jenkins JA, Walker WA, Schindler WG, Flores CH. An invitro evaluation of the accuracy of the Root ZX in the presence of various irrigants. *J Endodon.* 2001;27:209-11.
44. Ricard O, Roux D, Bourdeau L, Woda A. Clinical evaluation of the accuracy of the evident RCM Mark II apex locator. *J Endodon.* 1991;17:567-9.
45. Kim E, Lee SJ. Electronic apex locator. *Dental Clinics of North America.* 2004;48:35-54.
46. Fouad AF, Krell KV, McKendry DJ, Koorbusch GF, Olson RA. A clinical evaluation of five electronic root-canal length measuring instruments. *Journal of Endodontics.* 1990;16:446-9.
47. Jenkins JA, Walker WA, Schindler WG, Flores CM. An in vitro evaluation of the accuracy of the Root ZX in the presence of various irrigants. *J Endod.* 2001;27:209-11.
48. Vajrabhaya L, Tepmongkol P. Accuracy of apex locator. *Endod Dent Traumatol.* 1997;13:180-2.

Conflict of interest

The authors declare that they have no conflict of interest.